



PROGRAMMING MY VOICE 2.0

New wearers often find their own voice too loud when wearing hearing aids. My Voice 2.0 reduces return risks and helps with adapting to hearing aids by detecting when the member is speaking and improving the perception of their own voice — all without compromising prescribed gain for surrounding sounds.

1. At the initial fitting, instruct the member to wear the hearing aids and provide them with the instructions displayed on screen. Click Go to training.




2. Follow the instructions on screen and click Next.



3. Instruct the member to start counting after hearing a beep or when you point at them. Click Start training, count to 5 and then point at the wearer to continue.

When the training is complete, click close.



My Voice defaults to Med (medium) which works for most wearers.

If the member states that their voice is still too loud, you may move it to Max.

If the member states that their voice sounds weak or flat, you may move it to Min.

My Voice 2.0 makes it easy to manage the member's voice without compromising gain for surrounding sounds.

